


Name of Policy:	<u>Contingency Plan: Loss of Water/Code Copper</u>	 Effective Date: 9/1/2017 Initial Effective Date: 8/1/980
Policy Number:	3364-134-07	
Department:	Radiation Oncology	
Approving Officer:	Chief Executive Officer - UTMC Professor & Chairman, Radiation Oncology	
Responsible Agent:	Technical Manager, Radiation Oncology	
Scope:	Radiation Oncology	
<input type="checkbox"/> New policy proposal <input type="checkbox"/> Major revision of existing policy		<input type="checkbox"/> Minor/technical revision of existing policy <input checked="" type="checkbox"/> Reaffirmation of existing policy

(A) Policy Statement

The following procedure is established to assure that standards of safe practice are followed by all personnel in response to a loss of water in the department.



(B) Purpose of Policy

The purpose is to provide guidelines to assure that the area is maintained in a safe manner.

(C) Procedure

1. In the event of loss of water notify Facilities Maintenance at (419) 383-5353 or (419) 383-4298.
2. Notify the Manager and Physics staff.
3. When water has been restored to the department, the linear accelerators will be brought back up on line following established procedures and calibration checks.

Reference Hospital Policy # EP-08-014

Approved by:  _____ Changhu Chen, MD Professor & Chairman, Radiation Oncology	Review/Revision Date: 8/1995 4/1999 10/1986 3/2002 1987 1/2007 1988 10/15/2008 1989 7/1/2011 2/1990 9/1/2014 10/1992 9/1/2017 1/1994 3/1994 2/1995 1996 10/1996
 _____ Daniel Barbee, RN, BSN, MBA Chief Executive Officer - UTMC	Next Review Date: 9/1/2020
Review/Revision Completed By: Michelle Giovanoli	
Policies Superseded by This Policy: 38-12B	

It is the responsibility of the reader to verify with the responsible agent that this is the most current version of the policy.