



THE UNIVERSITY OF TOLEDO
MEDICAL CENTER

**LIFEPAK® 20 Defibrillator/Monitor
OPERATOR'S CHECKLIST**
For 2005 AHA/ERC Guidelines consistent devices
(cprMAX™ Technology)

Unit _____ Defib Biomed# _____ Year _____

	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>	OK <input checked="" type="checkbox"/>
<p align="center"><u>DEFIBRILLATORS</u></p> <p>1.) Check auto test printed report for SELF TEST SUCCEEDED. The therapy cable must be plugged into the QUIK-COMBO test plug for this feature to work.</p> <p align="center">***If self test succeeded and move onto #3 on back of page.</p>							
<p>2.) If auto test prints SELF TEST FAILED when plugged in AC power the service LED will illuminate. A MANUAL TEST MUST BE COMPLETED. MAKE SURE THAT QUIK-COMBO TEST PLUG PLUGGED IN and machine connected into AC power.</p> <p align="center">***If self test failed, complete the following steps and continue steps 3 through 9 on the back of page.</p> <p><u>For QUIK-COMBO® therapy cable:</u></p> <ul style="list-style-type: none"> • Confirm test plug connected to therapy cable • Press ANALYZE button. • After <i>ANALYZING NOW</i> message, look for the <i>REMOVE TEST PLUG</i>. • Connect QUIK-COMBO therapy cable to the energy reservoir • Open door and select ENERGY select • Select 10J and press CHARGE button • When fully charged press SHOCK button and look for the <i>ENERGY</i> 							
<p><u>For HARD PADDLES:</u></p> <ul style="list-style-type: none"> ▪ Confirm paddles properly seated in wells ▪ Select 10J and press the CHARGE button ▪ When fully charged press the SHOCK buttons on paddles, and look for <i>ENERGY</i> 							

3.) Inspect physical condition for: Foreign substances Clean the defibrillator.							
4.) Inspect power source for: AC power connector plugged into unit and AC power source; AC Mains LED is lit.							
5.) Examine accessory cables for: Cracking, damage, broken or bent parts or pins, and paddle surfaces for pitting.							
6.) Disconnect the defibrillator from AC power, wait 2 seconds, press ON and look for: Momentary <i>SELF-TEST</i> messages and momentary illumination of LEDs. Service LED is lit If lit, power off device, then power back on. If service LED stays lit contact Biomed. <i>LOW BATTERY/CONNECT TO AC POWER</i> Message connect to AC power to charge battery							
7.) Check ECG printer for: Adequate paper supply Ability to print							
8.) Confirm therapy cable connected to defibrillator and defibrillator disconnected from AC power: <u>For QUIK-COMBO® therapy cable:</u> <ul style="list-style-type: none"> • Confirm test plug connected to therapy cable • Press ANALYZE button. • After <i>ANALYZING NOW</i> message, look for the <i>REMOVE TEST PLUG</i>. • Connect QUIK-COMBO therapy cable to the energy reservoir • Open door and select ENERGY select • Select 10J and press CHARGE button • When fully charged press SHOCK button and look for the <i>ENERGY</i> 							
<u>For HARD PADDLES:</u> <ul style="list-style-type: none"> ▪ Confirm paddles properly seated in wells ▪ Select 10J and press the CHARGE button ▪ When fully charged press the SHOCK buttons on ▪ paddles, and look for <i>ENERGY</i> 							
9.) Reconnect the defibrillator to AC power and power off device.							
DATE							
INITIAL							

*Any problems with equipment call Biomedical Engineering, 419-383-4899