


Name of Policy: Handling Resert and other Disinfectants Policy Number: 3364-106-E17 Approving Officer: Chief Executive Officer-UTMC Responsible Agent: Director of Cardiovascular Services Scope: University of Toledo Medical Center		 Effective date: 1/8/2025 Original effective date: 5/2011	
Key words: Resert, Echo, Disinfectant, Safety Precautions, Instructions for Use and Disposal			
<input type="checkbox"/>	New policy proposal	<input type="checkbox"/>	Minor/technical revision of existing policy
<input type="checkbox"/>	Major revision of existing policy	<input checked="" type="checkbox"/>	Reaffirmation of existing policy

(A) Policy Statement

Personnel will follow appropriate safety precautions when using, mixing, or disposing of Resert and other disinfectants.

(B) Purpose of Policy

To guide the use of Resert in the Heart Station.

(C) Procedure

1. Personnel will wear protective attire, mask, gloves, long sleeved gown, and protective eyewear when mixing and disposing of disinfectants.
2. Resert will be monitored for concentration levels when bottle is opened and before every use. Personnel will follow manufacturer’s guidelines.
3. All containers with disinfectants will be clearly labeled with date opened, time opened, and expiration date.

<p>Approved by:</p> <p>/s/</p> <hr/> <p>Todd Korzec, RN, BSN Director, Cardiovascular Services</p> <p>1/8/2025</p> <hr/> <p>Date</p> <p>/s/</p> <hr/> <p>Samer J. Khouri MD Medical Director, Non-Invasive Cardiac Imaging</p> <p>1/8/2025</p> <hr/> <p>Date</p> <p>/s/</p> <hr/> <p>Christine Stesney-Ridenour, FACHE Chief Operating Officer</p> <p>1/8/2025</p> <hr/> <p>Date</p> <p><i>Review/Revision Completed by: Director, Cardiovascular Services</i></p>	<p>Policies Superseded by This Policy:</p> <ul style="list-style-type: none">• <i>None</i> <p>Initial effective date: 5/2011</p> <p>Review/Revision Date:</p> <p>5/25/2011 8/1/2011 6/2013 3/2016 9/2018 9/2021 3/2022 1/8/2025</p> <p>Next review date: 1/8/2028</p>
---	--