


<b>Name of Policy:</b> <u>EEG Monitoring during Tilt Table Procedure</u> <b>Policy Number:</b> 3364-138-10 <b>Department:</b> Neurodiagnostics <b>Approving Officer:</b> Associate VP of Patient Care Services UTMC <b>Responsible Agent:</b> Manager, Neurodiagnostics <b>Scope:</b> The University of Toledo Medical Center Neurodiagnostics	  <b>Effective Date: 6/1/2020</b> Initial Effective Date: 3/10/2004
<input type="checkbox"/> New policy proposal <input type="checkbox"/> Major revision of existing policy	<input type="checkbox"/> Minor/technical revision of existing policy <input checked="" type="checkbox"/> Reaffirmation of existing policy

Purpose of procedure is to provide guidelines for performing an EEG during a tilt table procedure.

1. Scheduling :
  - a. Test time will be coordinated with heart station.
  - b. Patient will be in EEG 1 ½ hours prior to the tilt table testing time.
2. EEG set up:
  - a. Electrodes are applied according to Electrode Application & Removal procedure, #138-06, using collodion or adhesive tape.
  - b. Head should be wrapped.
3. Recording:
  - a. Baseline should be performed if an EEG has not been performed in the past 90 days.
  - b. Prior to table being raised we should record for 5-10 minutes with video.
  - c. Baseline BP and Pulse should be documented on recording. All subsequent blood pressure, pulse should also be documented.
  - d. Documentation on record will include any positional changes, any patient symptoms and any injections given. This will include medication given and amount.
4. Ending test:
  - a. EEG should be continued for a minimum of 5 minutes past conclusion of tilt table.
  - b. Remove electrodes and complete all necessary paperwork.

<b>Approved by:</b>  /s/ _____ Cynthia Zapotosky, BSN, RN Manager, Neurodiagnostics Date _____  /s/ _____ Moneeca Smith, MSN, RN Associate VP of Patient Care Services Date _____	<b>Review/Revision Date:</b> 09/13/2004 10/03/2006 06/29/2007 08/11/2010 08/12/2011 06/01/2017 5/29/2020
	<b>Next Review Date:</b> 06/01/2023
<b>Policies Superseded by This Policy:</b> 17-1-10	