

NURSING SERVICE GUIDELINES HEMODIALYSIS

Guideline: **Dating Dialysate Jugs**

Policy Number Superseded:



Responsibility: Hemodialysis Registered Nurse

Effective Date:
June 2024

Purpose of Guidelines: To ensure dialysate jugs are properly labeled.

Initial Effective Date:
May 2015

Procedure	Point of Emphasis
All opened dialysate jugs must be labeled for discard 24 hours after opening.	Discard date and time will be visible on front of jug.
Jugs cannot be combined.	
Any jug remaining in cabinet on expiration date will be discarded.	

Approved by:

*Kurt Kless, MSN, MBA, RN, NE-BC
Chief Nursing Officer*

Initial effective date:

May 2015

Review/Revision Completed by:

Paula Kosmider, BSN, RN

Review/Revision Date:

May 2018

April 2021

June 2024

Next review date:

June 2027