


Name of Policy: Dry-Run on EDGE linear accelerator Policy Number: 3364-134-105 Approving Office: Chief Executive Officer – UTMC Chair – Radiation Oncology - UTMC Responsible Agent: Technical Manager, Radiation Oncology Scope: Radiation Oncology		 Effective Date: 7/1/2022 Initial Effective Date: 5/2015	
	New policy proposal		Minor/technical revision of existing policy
	Major revision of existing policy	X	Reaffirmation of existing policy

(A) Policy statement

A dry run will be performed on all patients pre-treatment on the EDGE linear accelerator.

(B) Purpose of policy

To ensure that the machine in which the treatment planned for the patient clears the patient, and table. Guarantee that no collisions will occur.

(C) Procedure

- Set the patient up on the treatment table.
- Make sure that the CBCT clears prior imaging.
- Check all angles of planned treatment fields.
- If patients' treatment plan is checked and all clearance is confirmed, therapists may rotate the table from outside the treatment room.

Approved by: _____ /s/ Mersiha Hadziahmetovic, MD Associate Professor & Chair Radiation Oncology _____ /s/ Richard P Swaine Chief Executive Officer - UTMC Review/Revision Completed By: Michelle Giovanoli		Review/Revision Date: 5/2015 6/1/2019 7/1/2022 Next Review Date: 7/1/2025
07/08/2022 Date	07/08/2022 Date	